

Nitrite HR No.1 TC

UFI: C9F9-Y53N-2M1V-KF01



Danger

H314 - Causes severe skin burns and eye damage

P260 - Do not breathe dust, fume, gas, mist, vapors and spray

P280 - Wear protective gloves/protective clothing and eye/face protection

P303 + P361 + P353 - IF ON SKIN (or hair): Take off immediately all

contaminated clothing. Rinse skin with water or shower

P305 + P351 + P338 - IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing

P310 - Immediately call a POISON CENTER or doctor

Nitrite HR No.2 TC

UFI: 96C5-7R41-DM1H-8388

H412 - Harmful to aquatic life with long lasting effects

P273 - Avoid release to the environment

P501 - Dispose of contents and container in accordance with local, regional, national, and international regulations as applicable

Water-i.d. GmbH
Daimlerstr. 20
D-76344 Eggenstein
Tel. +49 (0) 721 - 782029-0
Fax +49 (0) 721 - 782029-11
www.water-id.com
info@water-id.com

Test Kit

NaNO₂ (70 - 2800 mg/l)



Nitrite HR No. 1 TC



Nitrite HR No. 2 TC



Nitrite Tablet-count method



Nitrit Tabletten-Zählverfahren



Nitrite Methode de comptage à pilule



Nitrito Metodo di conteggio de pastiglie



Nitrito Método Contador de tabletas



Nitrit Tablettællings-metoden




Nitriet Tablet-telmethode

Range: 70 - 700 mg/l NaNO_2 : 20 ml sample water + 30 ml distilled water!
 Range: 140 - 1400 mg/l NaNO_2 : 10 ml sample water + 40 ml distilled water!
 Range: 280 - 2800 mg/l NaNO_2 : 5 ml sample water + 45 ml distilled water!




1



Fill in sample water!
(20 / 10 / 5 ml)

2




Add distilled water!
(30 / 40 / 45 ml)

3



2x Add Nitrite HR No. 1 TC*

4



5




1x
Add Nitrite HR No.2

6



7



Range 70 - 700 mg/l: Nitrite No. 2 x 70 = mg/l NaNO_2

Range 140 - 1400 mg/l: Nitrite No. 2 x 140 = mg/l NaNO_2

Range 280 - 2800 mg/l: Nitrite No. 2 x 280 = mg/l NaNO_2

