

Sulfite No.1 TC

UFI: PWWK-3895-F02E-HNGS



Warning

H373 - May cause damage to organs through prolonged exposure if swallowed

P260 - Do not breathe dust/fume/gas/mist/vapours/spray

P314 - Get medical advice/attention if you feel unwell

P501 - Dispose of contents/ container to an approved waste disposal plant

Test Kit

Na₂SO₃ (10 - 500 mg/l)

(High Range)



Sulphite No. 1 TC



Sulphite No. 2 HR TC

Tablet-count method

Tabletten-Zählverfahren

Méthode de comptage à pilule

Metodo di conteggio de pastiglie

Método Contador de tabletas

Tablettællings-metoden

Tablet-telmethode

NOTRUF/EMERGENCY: +49 (0)89-19240 (24h)

Water-i.d. GmbH
Daimlerstr. 20
D-76344 Eggenstein
Tel. +49 (0) 721 - 782029-0
Fax +49 (0) 721 - 782029-11
www.water-id.com
info@water-id.com



Sulphite



Sulfit



Sulfite



Sulfito




Sulfito




Sulfit




Sulfiet



1




2



3



4



5



50 ml: (Sulphite No. 2 HR x 40) - 20 = mg/l Na_2SO_3

100 ml: (Sulphite No. 2 HR x 20) - 10 = mg/l Na_2SO_3

